




M-6281 2



CORD CLAMPS ACCOMMODATE THE FOLLOWING:  
8013-V - .300-.360 DIA. INSULATION  
8015-V - .360-.575 DIA. INSULATION



| LIST OF PARTS     |          |          |
|-------------------|----------|----------|
| DESCRIPTION       | MATERIAL | FINISH   |
| HOUSING           | NYLON    | WHITE    |
| CONTACT BLADES    | BRASS    | 1-WHITE  |
| GR. CONTACT BLADE | BRASS    |          |
| CORD CLAMP SCR.   | STEEL    | ZINC PL. |
|                   |          |          |
|                   |          |          |
|                   |          |          |
|                   |          |          |
|                   |          |          |
|                   |          |          |
|                   |          |          |
|                   |          |          |
|                   |          |          |
|                   |          |          |
|                   |          |          |

| SYM | REVISIONS   | APP | DATE              |
|-----|---|-----|-------------------|
| 2   | ADD 25 SYMBOL PER<br>DIN # 1553                                 | CB  | 11/20/92          |
| 1   | DELETE FOR 3-WIRE, BRIDGEPORT<br>HUBBELL MARKING HOSPITAL GRADE | CB  | 9-28-92<br>A.J.L. |

DIMENSION SHEET FOR CAT. NO. 8013-V & 8015-V  
 REPAIRABLE  NON-REPAIRABLE  B M-6281 2

TITLE  
HOSPITAL GRADE 15A. 125V.  
3-WIRE DEAD FRONT VALISE  
PLUG WITH INSULATION  
DISPLACEMENT TERMINALS

TOLERANCES UNLESS OTHERWISE SPECIFIED  
 FRACTIONS ± 1/64  
 DECIMALS ± .005  
 ANGLES ± 2°

HARVEY HUBBELL, INC.  
 BRIDGEPORT, CONN.  
 DR. BY A.J.L. APP. BY D.L.  
 TR. BY ..... SCALE FULL  
 CHK'D. BY ..... DATE 1-29-82